

Test Certificate

CERTIFICATE No: **TRA-037751-33-20A**

ISSUE: **0**

DATE: **2020-05-06**

PURPOSE OF TEST: **Ingress Protection IP65**

CLIENT: **Opgal Optronic Industries Ltd.**

EQUIPMENT UNDER TEST: **EyeCGas 2.0 Optical Gas Imager**

TEST SPECIFICATIONS: **IEC 60529:1989+AMD1:1999+AMD2:2013
EN 60529:1992+A2:2013**

TEST DATES: **IP6X: 2019-05-06 (2 samples tested)
IPX5: 2019-05-14 (2 samples tested)
NOTE: A total of 4 separate samples of EyeCGas 2.0
were tested**

TEST LOCATION: **SKELMERSDALE**

TESTS CARRIED OUT: **IPX5 - Test clauses 14
IP6X - Test clauses 12 and 13**

TEST RESULTS: **The equipment met the requirements of the test
standard for an IP65 rating**

WRITTEN BY:
A Norris



TITLE:
ExTL Test Technician

APPROVED BY:
E Gadsby



TITLE:
Operations Manager - ATEX

The results herein relate only to the particular samples of equipment tested and the specific tests performed, as detailed above, and in accordance with the contract. Full details of test results, modifications and marginal results are held by Element Materials Technology Warwick Ltd. The quality control arrangements are in accordance with our UKAS accreditation. No representation or warranty is given that the tests performed under the terms of contract constitute, in themselves, a sufficient programme for the client's purpose, nor that the client's equipment is suitable for any particular purpose, nor that any approval has or will be granted by Element Materials Technology Warwick Ltd or any other body. The contents of this certificate shall not be reproduced, except in full, without the written approval of Element Materials Technology Warwick Ltd.

EMTEACC02